DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center – WO66-G609
Silver Spring, MD 20993-0002

May 12, 2015

ZIBO SANYING TRADE CO., LTD
C/O Mr. Ray Wang
Official Correspondent
Beijing Believe Tech. Service Co., Ltd.
1-202, Build 3, Beijing New World,
No. 5 Chaoyang Rd., Chaoyang District,
Beijing, 100024
CHINA

Re: K150224
Trade/Device Name: Blue Vinyl Examination Gloves Powder Free
Regulation Number: 21 CFR 880.6250
Regulation Name: Vinyl Patient Examination Gloves
Regulatory Class: I
Product Code: LYZ
Dated: March 29, 2015
Received: April 1, 2015

Dear Mr. Wang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm.

Sincerely yours,

Tejashri Purohit-Sheth, M.D.

Tejashri Purohit-Sheth, M.D.
Clinical Deputy Director
DAGRID/ODE/CDRH FOR

Erin I. Keith, M.S.
Division Director
Division of Anesthesiology, General Hospital,
Respiratory, Infection Control and
Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Section II Indications for Use

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: January 31, 2017
See PRA Statement below.

510(k) Number *(if known)*
K150224

Device Name
BLUE VINYL EXAMINATION GLOVES POWDER FREE

Indications for Use *(Describe)*

The BLUE VINYL EXAMINATION GLOVES POWDER FREE is a disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.

Type of Use *(Select one or both, as applicable)*

☐ Prescription Use (Part 21 CFR 801 Subpart D) ☒ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

Section III 510(k) Summary

This 510(k) Summary of 510(k) safety and effectiveness information is being submitted in accordance with requirements of SMDA 1990 and 21 CFR 807.92.

The assigned 510(k) Number: K150224

1. Date of Preparation: 2015/1/30

2. Sponsor

ZIBO SANYING TRADE CO., LTD
XILV VILLAGE FANGZHEN , ZHANGDIAN DISTRICT, ZIBO , SHANDONG , 255000 , CHINA

Contact Person: Ms. Jing Li
Tel: +86- 0533-3819144
Fax: +86-0533-3818127
Email: hclijing85@126.com

3. Submission Correspondent

Mr. Ray Wang
Beijing Believe Tech. Service Co., Ltd.
Tel: +86-21-50313932
Fax: +86-21-68093116
Email: Ray.Wang@believe-med.com

4. Proposed Device Identification

Trade Name: BLUE VINYL EXAMINATION GLOVES POWDER FREE
Device Name: Vinyl Patient Examination Gloves (Powder Free)
Common Name: Patient Examination Gloves

Classification: I
Product Code: LYZ
Regulation Number: 21 CFR 880.6250
Review Panel: General Hospital

Intended Use Statement:
The BLUE VINYL EXAMINATION GLOVES POWDER FREE is a disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.

5. Predicate Device Identification

510(k) Number: k142508
Product Name: Powder Free Yellow Vinyl Examination Gloves
Manufacturer: Zibo HUAQI TRADING Co., Ltd.

6. Device Description

The proposed device, BLUE VINYL EXAMINATION GLOVES POWDER FREE is a disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.

The proposed is BLUE VINYL EXAMINATION GLOVES POWDER FREE, and includes variations of different size.

7. Non-Clinical Test Conclusion

Bench tests were conducted to verify that the proposed device met all design specifications as was Substantially Equivalent (SE) to the predicate device. The test results demonstrated that the proposed device complies with the following standards:

ASTM D5250-06, Standard Specification for Poly(vinyl chloride) Gloves for Medical Application.
ASTM D 5151-06 (Reapproved 2011), Standard Test Method for Detection of Holes in Medical Gloves.
ASTM D6124-06 (Reaffirmation 2011), Standard Test Method for Residual Powder on Medical Gloves.
ISO 2859-1:1999, "Sampling Procedures for Inspection by Attributes – Part I: Sampling Plans Indexed by Acceptable Quality Level (AQL) for Lot-by-Lot Inspection.
ISO 10993-10: 2010, Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization.

8. Substantially Equivalent Comparison Conclusion

Table 1 General Comparison

ITEM	Proposed Device BLUE VINYL EXAMINATION GLOVES	Predicate Device (k142508) Powder Free Yellow Vinyl	Remark

	POWDER FREE	Patient Examination Gloves	
Product Code	LYZ	LYZ	SE
Regulation No.	21 CFR 880.6250	21 CFR 880.6250	SE
Class	I	I	SE
Intended Use	The BLUE VINYL EXAMINATION GLOVES POWDER FREE is a disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.	The Powder Free Yellow Vinyl Patient Examination Gloves is a disposable device intended for medical purposes that is worn on the examiner's hands or finger to prevent contamination between patient and examiner.	SE
Powdered or Powered free	Powdered free	Powdered free	SE

Table 2 Device Dimensions Comparison

	Designation	Size				Tolerance
		S	M	L	XL	
Proposed Device BLUE VINYL EXAMINATION GLOVES POWDER FREE	Length, mm	240	240	240	240	min
	Width, mm	85	95	105	115	±5
	Thickness, mm:					
	Finger	0.10				min
	Palm	0.08				min
	Cuff	0.06				min
	Designation	Size				Tolerance
		S	M	L	XL	
Predicate Device (k142508) Powder Free Yellow Vinyl Patient Examination Gloves	Length, mm	240	240	240	240	min
	Width, mm	85	95	105	115	±5
	Thickness, mm:					
	Finger	0.10				min
	Palm	0.08				min
	Cuff	0.06				min
Remark	SE					

Table 3 Performance Comparison

ITEM			Proposed Device BLUE VINYL EXAMINATION GLOVES POWDER FREE	Predicate Device (k142508) Powder Free Yellow Vinyl Patient Examination Gloves	Remark
Colorant			Blue	Yellow	Analysis
Physical Properties	Before Aging	Tensile Strength	13 MPa, min	13 MPa, min	SE
		Ultimate Elongation	400 % min	400 % min	SE
	After	Tensile	13 MPa, min	13 MPa, min	SE

	Aging	Strength			
		Ultimate Elongation	400 % min	400 % min	SE
	Comply with ASTM D5250			Comply with ASTM D5250	SE
Freedom from Holes			Be free from holes when tested in accordance with ASTM D5151	Be free from holes when tested in accordance with ASTM D5151	SE
Powder Content			0.50 mg per glove	Meet the requirements of ASTM 5250	SE

Table 4 Safety Comparison

ITEM		Proposed Device BLUE VINYL EXAMINATION GLOVES POWDER FREE	Predicate Device (k142508) Powder Free Yellow Vinyl Patient Examination Gloves	Remark
Material		Vinyl	Vinyl	SE
Biocompatibility	Irritation	Under the conditions of the study, not an irritant	Comply with ISO 10993-10	SE
	Sensitization	Under conditions of the study, not a sensitizer		
Label and Labeling		Meet FDA's Requirements	Meet FDA's Requirements	SE

The color of the proposed device is different from the color of the predicate device. This difference may potentially cause biocompatibility issues. In order to mitigate this risk, we conducted biocompatibility tests according to ISO 10993-10 and the test results demonstrated that the proposed device with blue colorant did not induce skin irritation and showed no significant evidence of causing skin sensitization.

The proposed device, BLUE VINYL EXAMINATION GLOVES POWDER FREE, is determined to be Substantially Equivalent (SE) to the predicate device in respect of safety and effectiveness.